



N95 Disposable Respirator
P/N: F04.1

INSTRUCTIONS FOR USE



DOC REF: IU-0045 - REV 1

MANUFACTURER DETAILS:

Air Filtration Solutions Ltd
72 Roman Way Industrial Estate,
Longridge Road, Preston, Lancashire
PR2 5BE. United Kingdom
Tel: 0044(0) 1772 790820
info@airfiltrationsolutions.co.uk
www.stealthmask.co.uk

APPROVALS:



42 CFR 84



FILTER LABEL SYMBOLS

See information supplied by manufacturer

Temperature range of storage conditions: -10°C to +40°C

Humidity range of storage conditions: <95%

Year/month Expiry Date

Instructions for use – N95 Disposable Respirator

The N95 Disposable Respirator is approved in accordance with NIOSH 42 CFR 84 as a NIOSH approved respirator for use in occupational applications.

The following table shows the product code and description used to identify the N95 Disposable Respirator:

Product Code	Product Description
F04.1	N95 Disposable Face Mask

GUIDELINES FOR THE SELECTION AND USE MAY VARY BETWEEN COUNTRIES. ALWAYS CHECK PRIOR TO USE TO ENSURE THAT YOUR SELECTION FOR THE INTENDED APPLICATION MEETS NATIONAL REQUIREMENTS.

Before occupational use of this respirator, a written respiratory protection program must be implemented meeting all the local government requirements. In the United States, employers must comply with OSHA 29 CFR 1910.134 which includes medical evaluation, training and fit testing.

APPLICATION

The N95 Disposable Respirator is suitable for the following applications:

Filter Class	Main Applications
N95 MINIMUM EFFICIENCY 95% @ 0.3 microns	Non-oil based: airborne particulates including pathogens Hazardous dust particles · Fumes

CHECKS PRIOR TO USE

- Prior to use ensure that the mask has been stored in accordance with the recommended storage instructions below and is still within its stated expiry date, the detail is displayed on the package label.
- Remove all packaging and inspect the mask for damage. If in doubt the suspected defective mask should be disposed of and replaced.
- Ensure that the intended application and atmosphere has been assessed by a competent person and meets the application limitations for the N95 Disposable Respirator (See application table above and the additional warnings and limitations).

FITTING INSTRUCTIONS

A: Ensure the checks prior to use above have been conducted before the use of each mask.

B: Remove eyewear. If facial hair protrudes beyond the border of the mask an effective seal will not be achieved and a powered respirator will be required.

C: Grasp the respirator with one hand positioning your thumb on the inside of the mask. (fig 1)

D: Pull the upper and lower headbands over the back of the head whilst positioning the respirator on your face. (fig 2)

E: Position the upper strap at the crown of your head and the lower strap at the neck. Ensure they are not twisted and are sitting flat. (fig 3)

F: Adjust the position of the respirator on the face and pinch the mask around the nose to form the metal strip ensuring a sealed face fit with no air leakage between the respirator and face. (fig 4)

G: Perform a positive pressure face fit check by exhaling deeply and feeling for air leakage around the nose and face to ensure an airtight seal has been achieved. Perform this each time the respirator is re-applied and repeat steps C – G if a leak is found. (fig 5)

USE

The N95 Disposable Respirator is suitable for **single shift use only**.

STORAGE

- The N95 Disposable Respirator should be stored in a dry and uncontaminated environment between temperatures of -10 to 40 degrees C.
- The N95 Disposable Respirator should not be stored for extended periods in direct sunlight or beyond the specified shelf life.
- If stored under conditions other than specified by the manufacturer the shelf life of the mask may be affected.

ADDITIONAL WARNINGS & LIMITATIONS

DO NOT use the N95 Disposable Respirator if:

- the air quality, the atmosphere or the type of hazard are unknown.
- the atmosphere is oxygen deficient (<19.5%).

- the atmosphere is oxygen enriched.
- the gas or vapour concentrations are unknown or exceed the value of the class of the mask.
- the atmosphere is explosive.
- the intended user has facial hair protruding beyond the border.

the mask has exceeded its expiry date, which is displayed on the mask pack label.

ALWAYS:

- Check the fit and effective seal of the facepiece prior to use – Note: it is unlikely that the requirement for leakage will be achieved if facial hair passes under the face seal.
- Check the respirator for damage and replace if in doubt.
- Ensure that the intended application and atmosphere has been assessed by a competent person and meets the application limitations.
- Check the expiry date, the detail is displayed on the pack label.

